

## **CERTIFICATE OF CALIBRATION**

Unit Model #:

195374

Unit Serial #:

LK360084G

Recorder Serial #: US27454-1-5-5009-PL1

Instrument #:

98503

MAC Address:

00:0A:8D:01:80:C7

Sure Controls does hereby certify the above instrument was calibrated against standards maintained by Sure Controls and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

## Displayed Value

| Recorder  | Primary<br>Standard<br>Signal<br>382.0 ° F | Primary<br>Standard<br>Signal<br>882.0° F | Primary<br>Standard<br>Signal<br>1382.0° F |
|-----------|--|---|--|
| CHANNEL 1 | 382  | 882                                       | 1382                                       |
| CHANNEL 2 | 382  | 882                                       | 1382                                       |
| CHANNEL 3 | 382  | 882                                       | 1382                                       |
| CHANNEL 4 | 382  | 882                                       | 1382                                       |
| CHANNEL 5 | 382  | 882                                       | 1382                                       |
| CHANNEL 6 | 382  | 882                                       | 1382                                       |

Calibration Date:

1/5/2010

Recalibration Date:

1/5/2011

Certified by:

VΗ

Instrumentation Used: Beamex Model No. MC5 Multi-Function Calibrator

Beamex Serial No. 25513615MJP

Sure Controls ID No. STD-035