

Miller ProHeat 35

Unit Serial Number: MC100320G Date Tested: 11/20/2013

Certified Calibration

Primary Standard						
	382° (F) 2.787VDC@RC9		882° (F) 6.013VDC@RC9		1382° (F) 9.239VDC@RC9	
	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display
TC1	2.783	382	6.010	882	9.231	1381
TC2	2.788	382	6.012	882	9.239	1382
TC3	2.793	383	6.016	883	9.241	1382
TC4	2.798	383	6.018	883	9.244	1383
TC5	2.791	383	6.005	883	9.241	1382
TC6	2.790	383	6.011	883	9.242	1383
Allowable	2.749-2.825	+/- 3° F	5.975-6.051	+/- 3° F	9.201-9.277	+/- 3° F

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 1882015 Calibration Date: 12/17/2012

Precision Digital Voltage Meter SN: 2036021 Calibration Date: 12/17/2012

Machine re-calibration due date: 11/20/2014

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.



Technician: Sal H. Date: 11/20/2013

Signature