



Sure Controls  
N981 Tower View Drive  
Greenville, WI 54942

## CERTIFICATE OF CALIBRATION

**Unit Model #:** 195374  
**Unit Serial #:** LG030051G  
**Recorder Serial #:** UR34889-001-010 0306 K28  
**Instrument #:** 29620  
**MAC Address:** 08:00:48:80:73:B4

Sure Controls does hereby certify the above instrument was calibrated against standards maintained by Sure Controls and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

| Recorder  | Displayed Value                     |                                     |                                      |
|-----------|-------------------------------------|-------------------------------------|--------------------------------------|
|           | Primary Standard Signal<br>382.0 °F | Primary Standard Signal<br>882.0 °F | Primary Standard Signal<br>1382.0 °F |
| CHANNEL 1 | 382                                 | 883                                 | 1383                                 |
| CHANNEL 2 | 382                                 | 883                                 | 1383                                 |
| CHANNEL 3 | 382                                 | 883                                 | 1384                                 |
| CHANNEL 4 | 382                                 | 883                                 | 1383                                 |
| CHANNEL 5 | 382                                 | 883                                 | 1383                                 |
| CHANNEL 6 | 382                                 | 883                                 | 1384                                 |

Calibration Date: 3/14/2012

Recalibration Date: 3/14/2013

Certified by: VH

Instrumentation Used: Beamex Model No. MC5 Multi-Function Calibrator  
Beamex Serial No. 25513615MJP  
Sure Controls ID No. STD-035