



Sure Controls
N981 Tower View Drive
Greenville, WI 54942

CERTIFICATE OF CALIBRATION

Unit Model #: 195374
Unit Serial #: LJ380172G
Recorder Serial #: US26221-1-10-1909-PL1
Instrument #: 60543
MAC Address: 00:0A:8D:00:EC:7F

Sure Controls does hereby certify the above instrument was calibrated against standards maintained by Sure Controls and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Recorder	Displayed Value		
	Primary Standard Signal 382.0 ° F	Primary Standard Signal 882.0° F	Primary Standard Signal 1382.0° F
CHANNEL 1	382	882	1382
CHANNEL 2	382	882	1382
CHANNEL 3	382	882	1383
CHANNEL 4	382	882	1382
CHANNEL 5	382	882	1382
CHANNEL 6	382	883	1383
CHANNEL 7	381	881	1380
CHANNEL 8	382	882	1382
CHANNEL 9	382	882	1382
CHANNEL 10	382	882	1382
CHANNEL 11	381	882	1381
CHANNEL 12	380	880	1380

Calibration Date: 7/13/2012

Recalibration Date: 7/13/2013

Certified by: SMP 

Instrumentation Used: Beamex Model No. MC5 Multi-Function Calibrator
Beamex Serial No. 25513615MJP
Sure Controls ID No. STD-035

SC1144 12/05