

Red-D-Arc Welderrentals

Calibration Certificate

Miller ProHeat 35

Unit Serial Number: LJ380313G

Date Tested: 5/22/2013

Certified Calibration

Primary Standard						
	382° (F) 2.787VDC@RC9		882° (F) 6.013VDC@RC9		1382° (F) 9.239VDC@RC9	
	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display
TC1	2.770	382	6.003	882	9.238	1382
TC2	2.776	382	6.008	882	9.240	1382
TC3	2.778	382	6.010	882	9.241	1382
TC4	2.779	382	6.008	882	9.237	1382
TC5	2.772	382	6.006	882	9.233	1382
TC6	2.773	382	6.004	882	9.234	1382
Allowable	2.749-2.825	+/- 3° F	5.975-6.051	+/- 3° F	9.201-9.277	+/- 3° F

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2083066

Calibration Date: 6/25/12

Precision Digital Voltage Meter SN: 2141018

Calibration Date: 8/15/12

Machine re-calibration due date: 5/22/2014

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.



Technician: DON WAGNER

Date: 5/22/2013

Signature