

Red-D-Arc  
685 Lee Industrial Blvd  
Austell GA 30168

## CERTIFICATE OF CALIBRATION

Unit Model Number: 195374  
Unit Serial Number: LJ390023G  
Calibration Date: April 12, 2012  
Re-Calibration Date: April 12, 2013

Certified By: TT  
Instrument Number: 60541  
MAC Address: 00:0A:8D:00:EC:7D

Arc-Heat LLC does hereby certify the above instrument was calibrated against standards maintained by Arc-Heat LLC and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

	PRIMARY STANDARD		PRIMARY STANDARD		PRIMARY STANDARD	
	Voltage In (DC)	Equivalent Temp (°F)	Voltage In (DC)	Equivalent Temp (°F)	Voltage In (DC)	Equivalent Temp (°F)
CHANNEL 1	2.787	382	6.013	882	9.239	1382
CHANNEL 2	2.787	382	6.013	882	9.239	1382
CHANNEL 3	2.787	382	6.013	881	9.239	1382
CHANNEL 4	2.787	382	6.013	882	9.239	1382
CHANNEL 5	2.787	382	6.013	882	9.239	1381
CHANNEL 6	2.787	382	6.013	882	9.239	1381

Instrumentation Used:

Calibrated Multimeter: HP/Agilent 34401A

Serial Number: US36049893