

# Red-D-Arc Welderrentals

## Calibration Certificate

Unit Serial Number: MA120026G

Date Tested: 27/09/13

Instrument Number: 95150

Mac Address: 00:0A:8D:01:73:AE

IP Address: 192.168.111.222

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	382	383	882	882	1382	1383
Channel 2	382	383	882	884	1382	1384
Channel 3	382	382	882	882	1382	1383
Channel 4	382	382	882	882	1382	1383
Channel 5	382	382	882	883	1382	1383
Channel 6	382	383	882	883	1382	1382

Allowable difference in temp reading is +or- 5 Deg. F.

**Instruments Used:**

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: 08/09/13

Precision digital multimeter SN: 2142011

Calibration Date: 08/09/13

ProHeat 35 power supply used for signal output SN: MA050188G

Calibration date: JULY/2013

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Calibration due date: 27/09/2014

Technician: Cuno Koch

Signature