

Red-D-Arc Welderrentals

Calibration Certificate

Miller ProHeat 35

Unit Serial Number: MA370104G

Date Tested: 6/21/13

Certified Calibration

Primary Standard						
	382° (F) 2.787VDC@RC9		882° (F) 6.013VDC@RC9		1382° (F) 9.239VDC@RC9	
	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display
TC1	2.787	382	6.010	882	9.235	1382
TC2	2.787	382	6.012	882	9.236	1382
TC3	2.788	382	6.010	882	9.231	1382
TC4	2.788	382	6.010	881	9.236	1382
TC5	2.786	381	6.009	881	9.232	1382
TC6	2.786	382	6.011	881	9.233	1382
Allowable	2.749-2.825	+/- 3° F	5.975-6.051	+/- 3° F	9.201-9.277	+/- 3° F

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2083066

Calibration Date: 6/25/12

Precision Digital Voltage Meter SN: 2141018

Calibration Date: 8/15/12

Machine re-calibration due date: 6/21/14

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.



Technician: Don Wagner

Date: 6/21/13

Signature