

# Red-D-Arc Welderrentals

## Calibration Certificate

Unit Serial Number: MC290371G  
 Instrument Number: 130831  
 IP Address: 192.168.111.222

Date Tested: 06/22/15  
 Mac Address: 00:0A:8D:01:FF:0F

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	383	383	883	883	1382	1384
Channel 2	383	383	883	884	1382	1384
Channel 3	383	383	882	884	1382	1383
Channel 4	383	383	882	884	1382	1385
Channel 5	383	383	882	882	1382	1383
Channel 6	383	382	882	883	1382	1384

Allowable difference in temp reading is +or- 5 Deg. F.

### Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: 24/07/14

Precision digital multimeter SN: 2142011

Calibration Date: 24/07/14

ProHeat 35 power supply used for signal output SN: LK230189G

Calibration date: 06/22/15

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Calibration due date: 06/22/16

Technician: Gino Rodin

Signature