

Red-D-Arc Welderrentals

Calibration Certificate

Miller ProHeat 35

Unit Serial Number: ME270376G

Date Tested: 8/5/16

Certified Calibration

Primary Standard						
	382° (F) 2.787VDC@RC9		882° (F) 6.013VDC@RC9		1382° (F) 9.239VDC@RC9	
	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display
TC1	2.796	382	6.021	882	9.248	1382
TC2	2.790	382	6.018	882	9.249	1382
TC3	2.790	382	6.019	882	9.250	1382
TC4	2.791	383	6.022	883	9.253	1383
TC5	2.800	383	6.027	883	9.251	1383
TC6	2.796	383	6.022	883	9.253	1383
Allowable	2.749-2.825	+/- 3° F	5.975-6.051	+/- 3° F	9.201-9.277	+/- 3° F

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 1882015

Calibration Date: 10/26/15

Precision Digital Voltage Meter SN: 2036021

Calibration Date: 10/26/15

Machine re-calibration due date: 8/5/2017

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.



Technician: Scott T.

Date: 8/5/2016

Signature